

Vibrating Table



- For the measurement of the Flow coefficient of aggregate.
- The device is designed to perform tests in accordance to EN933-6.
- The apparatus consists of a vibrating table mounted on a rugged base, a flow unit for coarse aggregates conforming to EN 933-6 and a sample of the reference material.
- Supplied with a certificate of conformity.
- Please note that the required voltage is 380V 3ph.
- Dimensions: 400mm x 400mm x 1400mm high